

TRIPLE LUMEN V-SET – *Instructions for Use*

REF: V3 GISS R



Store within 4°C - 30°C
to maintain shelf life



Keep away
from sunlight



Keep dry



Do not use if
package is damaged

STERILE R

Sterilized using irradiation



Non-pyrogenic



Do not re-use



NON-DEHP

CE
2797

LINE 1 (LARGE BORE PE LINED PVC)

– GRAVITY LINE WITH SPRING AND ANTI-REFLUX VALVE

Valve normally closed.

Opens with very low pressure 0.2 psi.

Anti-reflux valve prevents backflow for low incidence
of blockage / clotting.

LINES 2 and 3 (SMALL BORE PE LINED PVC)


– PUMP LINE / INJECTION PORT WITH ANTI-SIPHON VALVE

Valve normally closed.

Opens with high pressure 2.0 - 3.5 psi.

Do not connect drip sets to these lines.

Anti-siphon valve prevents backflow for low incidence of
blockage / clotting and inadvertent siphoning from cracked
syringe.



MRI COMPATIBLE

Go Medical Industries Pty Ltd

INNOVATIVE PRODUCTS TO IMPROVE COMMON MEDICAL PROCEDURES

LINE 1 (LARGE BORE PE LINED PVC)

– GRAVITY LINE WITH SPRING AND ANTI-REFLUX VALVE

Gravity infusion should be attached to the gravity line with spring and anti-reflux valve.

Recommended for Gravity and high infusion rates.

Total dead space 0.43 mL, common space = NIL

LINES 2 and 3 (SMALL BORE PE LINED PVC)

– PUMP LINE / INJECTION PORT WITH ANTI-SIPHON VALVE

Anti-siphon valves fitted for intermittent bolus injections or for infusions with electronic pumps.

Do not connect drip sets to these lines.

Total dead space 0.38 mL, common space = NIL

CONNECTION AND PRIMING

The V-set is designed to be connected directly to an IV cannula or catheter to eliminate common space. All fittings on the V-set are luer lock. All tubes are PE lined and therefore minimise drug sorption. Syringes should be connected to the lines without using needles. Use aseptic technique. Prime each line to remove air. Connect male luer lock fitting to IV cannula or catheter. Replace with a new set according to hospital infection control/intravenous protocol or within 96 hours.

If other fittings are connected to the IV cannula, the correct procedure to eliminate common space in the line is:

- 1) To obstruct the vein with pressure, remove the other fittings and insert the V-set directly into the cannula or
- 2) Not use the V-set due to the other fittings, but this will fail to allow 3 different access ports with no common space.

CAUTION



- Do not connect any valve or other fittings between the IV cannula or catheter and the V-set, as this will increase common space, thereby increasing the risk of flow rate errors (under-dosing), blockages, and/or inadvertent bolus delivery (over-dosing)
- Do not connect any valve or other fittings between the IV cannula or catheter and the V-set, as this may restrict flow rate or cause blockages.
- Do not change valve configuration, as this may increase the risk of backflow and blockages.
- Flow rate in small bore lines may be restricted and cause pressure alarms on high flow settings.
- Dispose of the device safely.



Go Medical Industries Pty Ltd

200 Churchill Avenue, Subiaco, WA 6008, Australia

Tel: +61 (0)8 9388 1700 Email: info@gomedical.com.au

Website: www.gomedical.com.au



Advena Limited

Tower Business Centre, 2nd Flr Tower Street, Swatar, BKR 4013, Malta